


Report Test

M/C : Henan Sanqiang Medical Equipment Liability Co. Ltd
Model : SQ-H220
Serial No : HNSQ2020050063

LOT No : 00974
Date : 14 May 2025

Determination of Sterilization Operation Specification

Item Detail	Specification	Setting	Result
Chamber Setting Temperature	40.0-55.5 °C	45.0 °C	
Vacuum pressure	40-60 kPa	60 kPa	
Vacuum Time	50-60 min	60 min	
Humidification			
Humidity	50.0-90.0%	20.2 %	
Sterilization			
Sterilization Temperature	40.0-55.5 °C	43.5 °C	
Gas Injection			
Final pressure	0-10 kPa	4 kPa	
Final temperature of vaporizer	40.0-55.5 °C	44.8 °C	
Exposure			
Temperature at start point of Exposure time period	40.0-55.5 °C	43.3 °C	
Temperature at finish point of Exposure time period	40.0-55.5 °C	46.0 °C	
Exposure Time	120-240 min	180 min	
Exhaust			
Exhaust Time	60-80 min	75 min	
First Vacuum pressure	0-2 kPa	1.5 kPa	
Second Vacuum pressure	0-2 kPa	1.5 kPa	
Third Vacuum pressure	0-2 kPa	1.5 kPa	
Fourth Vacuum pressure	0-2 kPa	1.5 kPa	
Fifth Vacuum pressure	0-2 kPa	1.5 kPa	
Storage			
Storage Time	≥ 48 hour	≥ 48 hour	
Chemical Test			
EO sterilization temperature 40-55°C, Time Sterilization by EO 3hours	Blue	Red	Pass
Biological Test			
EZ Test Biological Indicator	Green	Green	Pass


 (.....)
 Mr. Jutha Liam mukda
 Research and Development Engineer

			LOT	Color	BOX	TOTAL / PCS
		208935				
5	Z-JB-1.6-3-CR	STERILE SURGICAL STEEL PREMIUM CRYSTAL JEWELLED BALL	LOT NO 00974			5
		209425				
1	Z-XBL-7.1.6-16.0-6-CL	STERILE BIOFLEX BARBELL WITH UV ACRYLIC GLOW IN THE DARK BALL	LOT NO 00974			9
2	Z-XBLB-PINS-1.6-28.0-4-BK	STERILE BIOFLEX BARBELL WITH UV ACRYLIC GLOW IN THE DARK BALL	LOT NO 00974			10
3	Z-XBLPD-PINS-1.6-27.0-5-CL	STERILE BIOFLEX PLEASURE DOME ENDED BARBELL PIN	LOT NO 00974			10
4	Z-XINLBD-1.6-8.0-2-CL	STERILE BIOFLEX INTERNAL LABRET WITH DISC	LOT NO 00974			10
5	Z-XINLBINS-1.6-6.0-1.7X12-CL	STERILE BIOFLEX INTERNAL LABRET WITH SURGICAL STEEL INSERTION PIN	LOT NO 00974			2
		209438				
1	Z-BCSHJ112-1.2-8.0-AQ-CR	STERILE SURGICAL STEEL JEWELLED HINGED SEGMENT RING	LOT NO 00974			2
2	Z-BCSHJ114-1.2-8.0-CR	STERILE SURGICAL STEEL JEWELLED HINGED SEGMENT RING	LOT NO 00974			2
3	Z-ER003-6-RO	STERILE STERLING SILVER 925 JEWELLED PRONG SET ROUND EAR STUDS PAIR	LOT NO 00974			1
4	Z-ER006-5-CR	STERILE STERLING SILVER 925 JEWELLED PRONG SET HEART EAR STUDS PAIR	LOT NO 00974			1
5	Z-INMLBJ164-1.0-6.0-CR	STERILE SURGICAL STEEL INTERNALLY THREADED JEWELLED MICRO LABRET	LOT NO 00974			1
6	Z-SCSH225-1.2-10.0-CR-AQ	STERILE SURGICAL STEEL JEWELLED MULTI PURPOSE CLICKER	LOT NO 00974			2
		209441				
1	Z-MBLDJ-1.2-8.0-3-CR	STERILE SURGICAL STEEL DOUBLE JEWELLED MICRO BARBELL	LOT NO 00974			100
2	Z-MLBJ-1.2-10.0-3-CR	STERILE SURGICAL STEEL JEWELLED MICRO LABRET	LOT NO 00974			100
		209511				
1	Z-GPMLB-1.2-8.0-2-5	STERILE GOLD PVD COATED SURGICAL STEEL MICRO LABRET	LOT NO 00974			3
2	Z-GPTMBN-1.2-8.0-3	STERILE GOLD PVD COATED TITANIUM CURVED MICRO BARBELL	LOT NO 00974			3
3	Z-INMLBJ-1.2-10.0-2-CR	STERILE SURGICAL STEEL INTERNALLY THREADED PREMIUM CRYSTAL JEWELLED MICRO LABRET	LOT NO 00974			4
4	Z-INMLBJ-1.2-8.0-2-CR	STERILE SURGICAL STEEL INTERNALLY THREADED PREMIUM CRYSTAL JEWELLED MICRO LABRET	LOT NO 00974			5
5	Z-JNO-0.8-6.5-2.35-CR	STERILE SURGICAL STEEL PREMIUM CRYSTAL JEWELLED CURVED NOSE STUD	LOT NO 00974			3
6	Z-MBNDJ-1.2-8.0-3-CR	STERILE SURGICAL STEEL DOUBLE JEWELLED CURVED MICRO BARBELL	LOT NO 00974			4
7	Z-TJNO-0.8-6.5-2-CR	STERILE TITANIUM CURVED PRONG SET 2MM JEWELLED NOSE STUD	LOT NO 00974			4
		209512				
1	Z-GPTJNO-OP-0.8-6.5-2.35-CR	STERILE GOLD PVD COATED TITANIUM OPTIMA CRYSTAL JEWELLED CURVED NOSE STUD	LOT NO 00974			4
		209463				
45	Z-MLB-1.2-8.0-3	STERILE SURGICAL STEEL MICRO LABRET	LOT NO 00974			100
46	Z-TJNO-1.0-6.5-2.35-HP-CR	STERILE TITANIUM PREMIUM CRYSTAL JEWELLED CURVED NOSE STUD	LOT NO 00974			15
47	Z-TNO-1.0-6.5-2-HP	STERILE TITANIUM CURVED BALL NOSE STUD	LOT NO 00974			15
48	Z-TNO-1.0-6.5-2-5-HP	STERILE TITANIUM CURVED BALL NOSE STUD	LOT NO 00974			10
		209466				
1	Z-TMLB-1.2-8.0-3-HP	STERILE TITANIUM MICRO LABRET	LOT NO 00974			10
		209516				
2	Z-INMLBJD-1.2-10.0-2.5-CR	STERILE SURGICAL STEEL INTERNALLY THREADED LABRET WITH JEWELLED DISC	LOT NO 00974			16
3	Z-MBNDJ-1.2-8.0-3-CR	STERILE SURGICAL STEEL DOUBLE JEWELLED CURVED MICRO BARBELL	LOT NO 00974			8
4	Z-MLBJ-1.2-9.0-2.5-CR	STERILE SURGICAL STEEL JEWELLED MICRO LABRET	LOT NO 00974			4
		2541186				
1	PT-Z-GPTBL-1.6-20.0-5	STERILE GOLD PVD COATED TITANIUM BARBELL	LOT NO 00974			50
		2535774				
2	THD-Z-BKINSC427-1.2	STERILE BLACK PVD COATED SURGICAL STEEL MICRO ATTACHMENT FOR 1.2MM INTER	LOT NO 00974			300
		2540887				
2	Z-GPTMBN-1.2-9.0-3	STERILE GOLD PVD COATED TITANIUM CURVED MICRO BARBELL	LOT NO 00974			20
		2541351				
2	TDI-Z-TBL-1.6-16.0-5	STERILE TITANIUM BARBELL FOR TDI	LOT NO 00974			300
		2542383				
25	THD-Z-TINMCB-1.2-8.0-3	STERILE TITANIUM INTERNALLY THREADED MICRO CIRCULAR BARBELL	LOT NO 00974			100
		2542423				
20	PT-Z-TMBN-1.2-10.0-3	STERILE TITANIUM MICRO CURVED BARBELL FOR PTIT ROC	LOT NO 00974			200
26	PT-Z-TMLBJ-1.2-8.0-2.5-CR	STERILE TITANIUM MICRO LABRET WITH JEWELLED BALL FOR PTIT ROC	LOT NO 00974			500
		2538029				
6	SV-STERILIZE	SERVICE STERILE	LOT NO 00974			2
		2541131				
2	SV-STERILIZE	SERVICE STERILE	LOT NO 00974			20
		2542188				
2	SV-STERILIZE	SERVICE STERILE	LOT NO 00974			29
		2540931				
3	SV-STERILIZE	SERVICE STERILE	LOT NO 00974			20
		2543179				
2	SV-STERILIZE	SERVICE STERILE	LOT NO 00974			7
						2011

Sterilization:
Temperature: 44.7℃
Humidity: 20.1RH
Pressures: -3.6KPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2025-5-14 13:39

Add E0 gas
Temperature: 44.7℃
Humidity: 20.1RH
Pressures: -4.0KPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2025-5-14 13:39

Evacuation
Temperature: 44.3℃
Humidity: 20.1RH
Pressures: -0.0KPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2025-5-14 13:30

Keeping Temp
Temperature: 43.6℃
Humidity: 20.1RH
Pressures: -1.6KPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2025-5-14 12:30

Keep pressure qualified
Temperature: 42.4℃
Humidity: 20.1RH
Pressures: -56.7KPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2025-5-14 12:28

Evacuation
Temperature: 41.8℃
Humidity: 20.1RH
Pressures: -3.0KPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2025-5-14 12:17

Heating-up:
Temperature: 23.1℃
Humidity: 20.1RH
Pressures: -3.0KPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2025-5-14 11:40

Setting:
Temperature: 45.0℃
Humidity: 55.0RH
Pressures: -60.0KPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2025-5-14 11:40

1A/5/25

LOT 974

END

Air replacement:
Temperature: 45.1℃
Humidity: 20.2RH
Pressures: -1.6KPa
Sterilization time: 0Hour
Air replace rounds: 1
round time: 0min(s)
2025-5-14 18:6

Air replacement:
Temperature: 45.6℃
Humidity: 20.2RH
Pressures: -1.6KPa
Sterilization time: 0Hour
Air replace rounds: 2
round time: 0min(s)
2025-5-14 17:48

Air replacement:
Temperature: 45.5℃
Humidity: 20.2RH
Pressures: -1.6KPa
Sterilization time: 0Hour
Air replace rounds: 3
round time: 0min(s)
2025-5-14 17:25

Air replacement:
Temperature: 45.4℃
Humidity: 20.1RH
Pressures: -1.6KPa
Sterilization time: 0Hour
Air replace rounds: 4
round time: 0min(s)
2025-5-14 17:1

Air replacement:
Temperature: 45.4℃
Humidity: 20.3RH
Pressures: -1.6KPa
Sterilization time: 0Hour
Air replace rounds: 5
round time: 15min(s)
2025-5-14 16:45

Sterilization:
Temperature: 45.4℃
Humidity: 20.3RH
Pressures: -0.0KPa
Sterilization time: 0Hour
Air replace rounds: 5
round time: 1min(s)
2025-5-14 16:39

Sterilization:
Temperature: 44.9℃
Humidity: 20.3RH
Pressures: -3.9KPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 3min(s)
2025-5-14 13:41



