


Report Test

M/C : Henan Sanqiang Medical Equipment Liability Co. Ltd
Model : SQ-H220
Serial No : HNSQ2020050063

LOT No : 00975
Date : 15 May 2025

Determination of Sterilization Operation Specification

Item Detail	Specification	Setting	Result
Chamber Setting Temperature	40.0-55.5 °C	45.0 °C	
Vacuum pressure	40-60 kPa	60 kPa	
Vacuum Time	50-60 min	60 min	
Humidification			
Humidity	50.0-90.0%	20.2 %	
Sterilization			
Sterilization Temperature	40.0-55.5 °C	43.5 °C	
Gas Injection			
Final pressure	0-10 kPa	4 kPa	
Final temperature of vaporizer	40.0-55.5 °C	44.8 °C	
Exposure			
Temperature at start point of Exposure time period	40.0-55.5 °C	43.3 °C	
Temperature at finish point of Exposure time period	40.0-55.5 °C	46.0 °C	
Exposure Time	120-240 min	180 min	
Exhaust			
Exhaust Time	60-80 min	75 min	
First Vacuum pressure	0-2 kPa	1.5 kPa	
Second Vacuum pressure	0-2 kPa	1.5 kPa	
Third Vacuum pressure	0-2 kPa	1.5 kPa	
Fourth Vacuum pressure	0-2 kPa	1.5 kPa	
Fifth Vacuum pressure	0-2 kPa	1.5 kPa	
Storage			
Storage Time	≥ 48 hour	≥ 48 hour	
Chemical Test			
EO sterilization temperature 40-55°C, Time Sterilization by EO 3hours	Blue	Red	Pass
Biological Test			
EZ Test Biological Indicator	Green	Green	Pass


 (.....)
 Mr. Jutha Liam mukda

			LOT	Color	BOX	TOTAL / PCS
		209491				
1	Z-BKSCNC23-1.6-16.0-CR	STERILE BLACK PVD COATED SURGICAL STEEL JEWEL NIPPLE CLICKER	LOT NO 00975			2
		209515				
1	Z-XINLBINS-1.2-10.0-1.3X12-CL	STERILE BIOFLEX INTERNAL LABRET WITH SURGICAL STEEL INSERTION PIN	LOT NO 00975			2
		209542				
1	Z-XINDATJD-1.6-1.5-5-CL-HP-CR	STERILE BIOFLEX INTERNALLY THREADED DERMAL ANCHOR WITH TITANIUM JEWEL	LOT NO 00975			2
2	Z-XINDATJD-1.6-2.0-5-CL-HP-CR	STERILE BIOFLEX INTERNALLY THREADED DERMAL ANCHOR WITH TITANIUM JEWEL	LOT NO 00975			4
3	Z-XINDATJD-1.6-2.5-5-CL-HP-CR	STERILE BIOFLEX INTERNALLY THREADED DERMAL ANCHOR WITH TITANIUM JEWEL	LOT NO 00975			4
		209560				
1	Z-INMLBJD-1.2-10.0-2-CR	STERILE SURGICAL STEEL INTERNALLY THREADED LABRET WITH JEWEL DISC	LOT NO 00975			1
		209564				
1	Z-XINLBJS13-1.0-7.0-2-CL-CR	STERILE BIOFLEX INTERNAL LABRET WITH SILVER JEWEL ATTACHMENT	LOT NO 00975			2
		2538562				
1	Z-MELSJ1-1.2-10.0-3-CR	STERILE SURGICAL STEEL ROUND PRONG SET JEWEL TRAGUS MICRO BARBELL	LOT NO 00975			1000
		2542136				
1	Z-LJNO-1.0-8.0-2.35-CR	STERILE SURGICAL STEEL PREMIUM CRYSTAL JEWEL LARGE LEFT CURVE NOSE S	LOT NO 00975			100
		2542383				
4	THD-Z-BNDJ-1.6-10.0-5/8-BZ	STERILE SURGICAL STEEL DOUBLE JEWEL NAVE BANANA FOR THENDRO	LOT NO 00975			100
7	THD-Z-GPLB-1.6-10.0-4	STERILE GOLD PVD COATED SURGICAL STEEL LABRET FOR THENDRO	LOT NO 00975			100
11	THD-Z-JNO-1.0-6.5-2.35-AM	STERILE SURGICAL STEEL JEWEL CURVED NOSE STUD FOR THENDRO	LOT NO 00975			100
18	THD-Z-TINLB-1.6-10.0-4	STERILE TITANIUM INTERNALLY THREADED LABRET	LOT NO 00975			100
		2542445				
3	TDI-Z10TMLBJOP-1.2-10.0-3-CR	STERILE TITANIUM MICRO LABRET WITH OPTIMA CRYSTAL BALL	LOT NO 00975		SET/100	1000
		2542714				
1	TDI-Z-BL-1.6-16.0-5	STERILE SURGICAL STEEL BARBELL FOR TDI	LOT NO 00975			200
		2543277				
3	SV-STERILIZE	SERVICE STERILE	LOT NO 00975			10
4	SV-STERILIZE	SERVICE STERILE	LOT NO 00975			10
		2542188				
3	SV-STERILIZE	SERVICE STERILE	LOT NO 00975			20
		2543362				
2	SV-STERILIZE	SERVICE STERILE	LOT NO 00975			8
						2765

Sterilization:
Temperature: 43.3°C
Humidity: 20.1RH
Pressures: -0.0KPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2025-5-15 15:18

Add E0 gas
Temperature: 43.1°C
Humidity: 20.2RH
Pressures: -0.0KPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2025-5-15 15:19

Evacuation
Temperature: 44.4°C
Humidity: 20.2RH
Pressures: -1.7KPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2025-5-15 15:14

Keeping Temp
Temperature: 45.5°C
Humidity: 20.1RH
Pressures: -1.6KPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2025-5-15 14:14

Keep pressure qualified
Temperature: 45.3°C
Humidity: 20.1RH
Pressures: -50.0KPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2025-5-15 14:12

Evacuation
Temperature: 45.5°C
Humidity: 20.1RH
Pressures: -3.0KPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2025-5-15 14:0

Heating-up:
Temperature: 45.5°C
Humidity: 20.1RH
Pressures: -3.0KPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2025-5-15 14:0

Setting:
Temperature: 45.0°C
Humidity: 55.0RH
Pressures: -60.0KPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2025-5-15 14:0

15-5-15
END Lot 975

Air replacement:
Temperature: 45.1°C
Humidity: 20.7RH
Pressures: -1.0KPa
Sterilization time: 0Hour
Air replace rounds: 1
round time: 0min(s)
2025-5-15 13:45

Air replacement:
Temperature: 45.8°C
Humidity: 20.1RH
Pressures: -1.0KPa
Sterilization time: 0Hour
Air replace rounds: 2
round time: 0min(s)
2025-5-15 13:25

Air replacement:
Temperature: 45.1°C
Humidity: 20.1RH
Pressures: -1.6KPa
Sterilization time: 0Hour
Air replace rounds: 5
round time: 0min(s)
2025-5-15 13:5

Air replacement:
Temperature: 45.1°C
Humidity: 20.1RH
Pressures: -1.0KPa
Sterilization time: 0Hour
Air replace rounds: 4
round time: 0min(s)
2025-5-15 13:45

Air replacement:
Temperature: 45.3°C
Humidity: 20.1RH
Pressures: -1.5KPa
Sterilization time: 0Hour
Air replace rounds: 6
round time: 15min(s)
2025-5-15 13:24

Sterilization:
Temperature: 44.4°C
Humidity: 20.2RH
Pressures: -0.0KPa
Sterilization time: 0Hour
Air replace rounds: 5
round time: 15min(s)
2025-5-15 13:10

Sterilization:
Temperature: 43.0°C
Humidity: 20.1RH
Pressures: -0.0KPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2025-5-15 15:20



FM-ISO0-009 Rev.02 (05/11/2020)