


## Report Test

M/C: Henan Sanqiang Medical Equipment Liability Co. Ltd  
 Model: SQ-H220  
 Serial No: HNSQ2020050063

LOT No: 001075  
 Date: 15 October 2025

### Determination of Sterilization Operation Specification

Item Detail	Specification	Setting	Result
Chamber Setting Temperature	40.0-55.5 °C	45.0 °C	
Vacuum pressure	40-60 kPa	60 kPa	
Vacuum Time	50-60 min	60 min	
<b>Humidification</b>			
Humidity	50.0-90.0%	20.2 %	
<b>Sterilization</b>			
Sterilization Temperature	40.0-55.5 °C	43.5 °C	
<b>Gas Injection</b>			
Final pressure	0-10 kPa	4 kPa	
Final temperature of vaporizer	40.0-55.5 °C	44.8 °C	
<b>Exposure</b>			
Temperature at start point of Exposure time period	40.0-55.5 °C	43.3 °C	
Temperature at finish point of Exposure time period	40.0-55.5 °C	46.0 °C	
Exposure Time	120-240 min	180 min	
<b>Exhaust</b>			
Exhaust Time	60-80 min	75 min	
First Vacuum pressure	0-2 kPa	1.5 kPa	
Second Vacuum pressure	0-2 kPa	1.5 kPa	
Third Vacuum pressure	0-2 kPa	1.5 kPa	
Fourth Vacuum pressure	0-2 kPa	1.5 kPa	
Fifth Vacuum pressure	0-2 kPa	1.5 kPa	
<b>Storage</b>			
Storage Time	≥ 48 hour	≥ 48 hour	
<b>Chemical Test</b>			
EO sterilization temperature 40-55°C, Time Sterilization by EO 3hours	Blue	Red	Pass
<b>Biological Test</b>			
EZ Test Biological Indicator	Green	Green	Pass

  
 (.....)  
 Mr. Jutha Liammukda  
 Research and Development Engineer



			LOT	Color	BOX	TOTAL / PCS
		214994				
1	ZDAH-4.0	STERILE TWEEZERS FOR DERMAL ANCHOR OR SKIN DIVER	LOT NO 1075			5
2	Z.TLJNO-1.0-6.5-1.8-HP-CR	STERILE TITANIUM LARGE CURVE JEWELLED NOSE STUD	LOT NO 1075			12
3	Z-TMBL-1.2-16.0-3-HP	STERILE TITANIUM MICRO BARBELL	LOT NO 1075			10
		215018				
1	ZSNHP-1.2-50.0	STERILE SURGICAL STEEL INTERNALLY THREADED INSERTION PINS	LOT NO 1075			20
2	ZSST-C-0.9-1.2	STERILE SURGICAL STEEL INTERNALLY THREADED CURVED TAPER	LOT NO 1075			29
		215021				
1	Z-GPBCSH-0.6-6.0	STERILE GOLD PVD COATED SURGICAL STEEL HINGED SEGMENT RING	LOT NO 1075			6
2	Z-GPBCSH-0.8-9.0	STERILE GOLD PVD COATED SURGICAL STEEL HINGED SEGMENT RING	LOT NO 1075			4
3	Z-GPFSST-1.2-10.0	STERILE GOLD PVD COATED SURGICAL STEEL FAKE SEPTUM RING	LOT NO 1075			4
4	Z-INLBSJ-1.6-8.0-3-CR	STERILE SURGICAL STEEL INTERNALLY THREADED LABRET WITH PRONG SET ROUND JEWELLED ATTACHMENT	LOT NO 1075			8
5	Z-INMLBJ-1.6-8.0-2-CR	STERILE SURGICAL STEEL INTERNALLY THREADED PREMIUM CRYSTAL JEWELLED MICRO LABRET	LOT NO 1075			14
6	Z-ZGMBS-1.2-8.0-3-CR	STERILE ZIRCON GOLD PVD COATED SURGICAL STEEL JEWELLED FANCY CURVED MICRO BARBELL - TRIPLE GEM CURVE	LOT NO 1075			3
7	Z-ZGMLB-1.2-6.0-3	STERILE ZIRCON GOLD PVD COATED SURGICAL STEEL MICRO LABRET	LOT NO 1075			15
8	Z-ZGMLB-1.2-8.0-3	STERILE ZIRCON GOLD PVD COATED SURGICAL STEEL MICRO LABRET	LOT NO 1075			16
9	Z-ZGMLBJ-1.2-8.0-3-CR	STERILE ZIRCON GOLD PVD COATED SURGICAL STEEL JEWELLED MICRO LABRET	LOT NO 1075			6
		215034				
2	Z-GPBHJ-1.6-12.0-5/8-CR	STERILE GOLD PVD COATED SURGICAL STEEL DOUBLE PREMIUM CRYSTALS JEWELLED NAVEL BANANA	LOT NO 1075			30
3	Z-MBLSJ-1.2-8.0-3-CR	STERILE SURGICAL STEEL ROUND PRONG SET JEWELLED TRAGUS MICRO BARBELL	LOT NO 1075			150
		2554466				
7	THD-Z-GPINS-290-1.2	STERILE GOLD PVD COATED SURGICAL STEEL MICRO ATTACHMENT FOR 1.2MM INTERNALLY THREADED PINS	LOT NO 1075			400
		2555625				
1	QL-Z-GPDFSC-1.6-12.0-5/8-CR	STERILE GOLD PVD COATED SURGICAL STEEL ROUND PRONG SET 8MM CZ JEWELLED NAVEL BANANA	LOT NO 1075			30
		2557484				
5	TDI-Z-TBNDJ-OP-1.6-10.0-5/8-RO	STERILE TITANIUM JEWELLED NAVEL BANANA FOR TDI	LOT NO 1075			50
		2558314				
11	PT-1.0-TBNDJ-1.6-12.0-5/8-CR	STERILE TITANIUM JEWELLED NAVEL BANANA FOR PTIT ROC	LOT NO 1075		SET/50	500
		2559291				
1	Z-TNMLBJ-1.2-10.0-2-HP-CR	STERILE TITANIUM INTERNALLY THREADED JEWELLED MICRO LABRET	LOT NO 1075			15
		2557550				
8	SV-STERILIZE	SERVICE STERILE	LOT NO 1075			9
						1337



Sterilization:  
Temperature: 43.4°C  
Humidity: 20.1RH  
Pressures: -0.0KPa  
Sterilization time: 3Hour  
Air replace rounds: 5  
round time: 15min(s)  
2025-10-15 13:47

Add EO gas  
Temperature: 43.2°C  
Humidity: 20.1RH  
Pressures: -4.0KPa  
Sterilization time: 3Hour  
Air replace rounds: 5  
round time: 15min(s)  
2025-10-15 13:46

Evacuation  
Temperature: 44.2°C  
Humidity: 20.2RH  
Pressures: -0.0KPa  
Sterilization time: 3Hour  
Air replace rounds: 5  
round time: 15min(s)  
2025-10-15 13:41

Keeping Temp  
Temperature: 42.9°C  
Humidity: 20.1RH  
Pressures: -1.0KPa  
Sterilization time: 3Hour  
Air replace rounds: 5  
round time: 15min(s)  
2025-10-15 12:41

Keep pressure qualified  
Temperature: 41.5°C  
Humidity: 20.1RH  
Pressures: -58.7KPa  
Sterilization time: 3Hour  
Air replace rounds: 5  
round time: 15min(s)  
2025-10-15 12:39

Evacuation  
Temperature: 41.6°C  
Humidity: 20.2RH  
Pressures: -3.0KPa  
Sterilization time: 3Hour  
Air replace rounds: 5  
round time: 15min(s)  
2025-10-15 12:38

Heating-up:  
Temperature: 21.8°C  
Humidity: 20.1RH  
Pressures: -0.0KPa  
Sterilization time: 3Hour  
Air replace rounds: 5  
round time: 15min(s)  
2025-10-15 11:46

Setting:  
Temperature: 45.0°C  
Humidity: 55.0RH  
Pressures: -80.0KPa  
Sterilization time: 3Hour  
Air replace rounds: 5  
round time: 15min(s)  
2025-10-15 11:45

15/10/25  
END  
Lot 1025

Air replacement:  
Temperature: 45.9°C  
Humidity: 20.1RH  
Pressures: -1.0KPa  
Sterilization time: 3Hour  
Air replace rounds: 1  
round time: 0min(s)  
2025-10-15 18:15

Air replacement:  
Temperature: 45.7°C  
Humidity: 20.2RH  
Pressures: -1.0KPa  
Sterilization time: 3Hour  
Air replace rounds: 2  
round time: 0min(s)  
2025-10-15 17:55

Air replacement:  
Temperature: 45.7°C  
Humidity: 20.1RH  
Pressures: -1.0KPa  
Sterilization time: 3Hour  
Air replace rounds: 3  
round time: 0min(s)  
2025-10-15 17:34

Air replacement:  
Temperature: 45.2°C  
Humidity: 20.2RH  
Pressures: -1.0KPa  
Sterilization time: 3Hour  
Air replace rounds: 4  
round time: 0min(s)  
2025-10-15 17:14

Air replacement:  
Temperature: 44.8°C  
Humidity: 20.1RH  
Pressures: -1.0KPa  
Sterilization time: 3Hour  
Air replace rounds: 5  
round time: 15min(s)  
2025-10-15 16:53

Sterilization:  
Temperature: 44.5°C  
Humidity: 20.2RH  
Pressures: -0.0KPa  
Sterilization time: 3Hour  
Air replace rounds: 5  
round time: 15min(s)  
2025-10-15 16:47

Sterilization:  
Temperature: 44.1°C  
Humidity: 20.1RH  
Pressures: -0.0KPa  
Sterilization time: 3Hour  
Air replace rounds: 5  
round time: 15min(s)  
2025-10-15 13:49