


Report Test

M/C: Henan Sanqiang Medical Equipment Liability Co. Ltd
 Model: SQ-H220
 Serial No: HNSQ2020050063

LOT No: 001127
 Date: 10 January 2026

Determination of Sterilization Operation Specification

Item Detail	Specification	Setting	Result
Chamber Setting Temperature	40.0-55.5 °C	45.0 °C	
Vacuum pressure	40-60 kPa	60 kPa	
Vacuum Time	50-60 min	60 min	
Humidification			
Humidity	50.0-90.0%	20.2 %	
Sterilization			
Sterilization Temperature	40.0-55.5 °C	43.5 °C	
Gas Injection			
Final pressure	0-10 kPa	4 kPa	
Final temperature of vaporizer	40.0-55.5 °C	44.8 °C	
Exposure			
Temperature at start point of Exposure time period	40.0-55.5 °C	43.3 °C	
Temperature at finish point of Exposure time period	40.0-55.5 °C	46.0 °C	
Exposure Time	120-240 min	180 min	
Exhaust			
Exhaust Time	60-80 min	75 min	
First Vacuum pressure	0-2 kPa	1.5 kPa	
Second Vacuum pressure	0-2 kPa	1.5 kPa	
Third Vacuum pressure	0-2 kPa	1.5 kPa	
Fourth Vacuum pressure	0-2 kPa	1.5 kPa	
Fifth Vacuum pressure	0-2 kPa	1.5 kPa	
Storage			
Storage Time	≥ 48 hour	≥ 48 hour	
Chemical Test			
EO sterilization temperature 40-55°C, Time Sterilization by EO 3hours	Blue	Red	Pass
Biological Test			
EZ Test Biological Indicator	Green	Green	Pass


 (.....)
 Mr. Jutha Liammukda
 Research and Development Engineer

		LOT	QTY	BOX	TOTAL / PCS
		217490			
1	Z-MBL-1.2-10.0-3	STERILE SURGICAL STEEL MICRO BARBELL	LOT NO 001127		200
2	Z-MLB-1.2-12.0-3	STERILE SURGICAL STEEL MICRO LABRET	LOT NO 001127		100
3	Z-SES1-3-CR	STERILE SURGICAL STEEL ROUND PRONG SET JEWELLED EAR STUDS PAIR	LOT NO 001127		100
4	Z-SES1-4-CR	STERILE SURGICAL STEEL ROUND PRONG SET JEWELLED EAR STUDS PAIR	LOT NO 001127		150
5	Z-TMBL-1.2-10.0-3-HP	STERILE TITANIUM MICRO BARBELL	LOT NO 001127		50
		217513			
1	Z-SCES045	STERILE SURGICAL STEEL EAR STUDS PAIR - FLOWER	LOT NO 001127		6
2	Z-SCES077	STERILE SURGICAL STEEL EAR STUDS PAIR - HEART	LOT NO 001127		6
		217523			
2	Z-GPBNJ-1.5-10.0-5/8-CR	STERILE GOLD PVD COATED SURGICAL STEEL PREMIUM CRYSTAL JEWELLED NAVEL BANANA	LOT NO 001127		3
3	Z-GPBNJ-1.5-12.0-5/8-CR	STERILE GOLD PVD COATED SURGICAL STEEL PREMIUM CRYSTAL JEWELLED NAVEL BANANA	LOT NO 001127		3
4	Z-GPINMLBJ146-1.2-8.0-CR	STERILE GOLD PVD COATED SURGICAL STEEL INTERNALLY THREADED JEWELLED MICRO LABRET	LOT NO 001127		8
5	Z-GPJNO-0.8-6.5-1.8-CR	STERILE GOLD PVD COATED SURGICAL STEEL PREMIUM CRYSTAL JEWELLED CURVED NOSE STUD	LOT NO 001127		3
6	Z-GPJNO-1.0-6.5-2.35-CR	STERILE GOLD PVD COATED SURGICAL STEEL PREMIUM CRYSTAL JEWELLED CURVED NOSE STUD	LOT NO 001127		3
7	Z-GPSCRC022-1.2-8.0-CR	STERILE GOLD PVD COATED SURGICAL STEEL PRONG SET CRYSTAL JEWELLED ROOK CLICKER	LOT NO 001127		3
9	Z-JNO-3.6-6.5-1.8-CR	STERILE SURGICAL STEEL PREMIUM CRYSTAL JEWELLED CURVED NOSE STUD	LOT NO 001127		3
10	Z-JNO-0.8-6.5-2.35-CR	STERILE SURGICAL STEEL PREMIUM CRYSTAL JEWELLED CURVED NOSE STUD	LOT NO 001127		3
11	Z-MBN-1.0-6.0-2	STERILE SURGICAL STEEL CURVED MICRO BARBELL	LOT NO 001127		3
12	Z-MBN-1.0-8.0-2	STERILE SURGICAL STEEL CURVED MICRO BARBELL	LOT NO 001127		3
13	Z-SCRC007-1.2-8.0-CR	STERILE SURGICAL STEEL PRONG SET JEWELLED ROOK CLICKER	LOT NO 001127		3
14	Z-SCRC020-1.2-6.0	SURGICAL STEEL ROOK CLICKER - PLAIN SMOOTH	LOT NO 001127		3
15	Z-SCRC021-1.2-6.0-CR	STERILE SURGICAL STEEL PREMIUM CRYSTAL JEWELLED ROOK CLICKER	LOT NO 001127		3
16	Z-TMBL-1.2-38.0-3-HP	STERILE TITANIUM MICRO BARBELL	LOT NO 001127		3
17	Z-XBN-1.1-6.10-0-4-CL	STERILE BIOFLEX CURVED BARBELL WITH STEEL BALLS	LOT NO 001127		3
18	Z-ZGTMBN-1.2-8.0-3	STERILE ZIRCON GOLD PVD COATED TITANIUM CURVED MICRO BARBELL	LOT NO 001127		3
19	Z-ZGCNO-0.8-6.5-2	STERILE ZIRCON GOLD PVD COATED SURGICAL STEEL CURVED BALL NOSE STUD	LOT NO 001127		3
		255486			
9	THD-Z-GPTINC-1.6-4	STERILE GOLD PVD COATED TITANIUM CONE FOR 1.6MM INTERNALLY THREADED PINS	LOT NO 001127		400
		2563413			
69	THD-Z-TINMLBJ-1.2-10.0-3-YE-CR	STERILE ANODISED TITANIUM INTERNALLY THREADED JEWELLED MICRO LABRET	LOT NO 001127		130
		2568223			
1	Z-SCBCSH 2.5	STERILE SURGICAL STEEL HINGED SEGMENT RING	LOT NO 001127		19
		2568583			
1	OLA-Z-JNO-1.0-8.0-2.35-CR	STERILE SURGICAL STEEL PREMIUM CRYSTAL JEWELLED CURVE NOSE STUD	LOT NO 001127		100
		2568458			
3	PT-Z-GPTBCSH-1.2-8.0	STERILE GOLD PVD COATED TITANIUM HINGED SEGMENT RING	LOT NO 001127		150
7	PT-Z-GPTMBL-1.2-10.0-3	STERILE GOLD PVD COATED TITANIUM MICRO BARBELL	LOT NO 001127		50
8	PT-Z-GPTMBN-1.2-10.0-3	STERILE GOLD PVD COATED TITANIUM CURVED MICRO BARBELL	LOT NO 001127		150
9	PT-Z-GPTMCB-1.2-10.0-3	STERILE GOLD PVD COATED TITANIUM MICRO CIRCULAR BARBELL	LOT NO 001127		100
10	PT-Z-GPTMLB-1.2-10.0-3	STERILE GOLD PVD COATED TITANIUM MICRO LABRET	LOT NO 001127		400
11	PT-Z-GPTMLB-1.2-8.0-3	STERILE GOLD PVD COATED TITANIUM MICRO LABRET	LOT NO 001127		1000
22	PT-Z-TINMBN-1.2-8.0-3-HP	STERILE TITANIUM INTERNALLY THREADED CURVED MICRO BARBELL FOR PTTT ROC	LOT NO 001127		50
26	PT-Z-TMBN-1.2-10.0-3	STERILE TITANIUM MICRO CURVED BARBELL FOR PTTT ROC	LOT NO 001127		150
28	PT-Z-TMCB-1.2-10.0-3	STERILE TITANIUM MICRO CIRCULAR BARBELL FOR PTTT ROC	LOT NO 001127		100
32	PT-Z-NINNOTJ02-0.8-6.6-0-2-CL-CR	STERILE BIOFLEX INTERNAL CURVED NOSE STUD WITH TITANIUM JEWELLED DISC FOR PTTT ROC	LOT NO 001127		50
37	PT-Z-10-TBL-1.6-22.0-5	STERILIZED TITANIUM BARBELL FOR PTTT ROC	LOT NO 001127	SET/15	150
40	PT-Z-10-TMBN-1.2-8.0-3	STERILE TITANIUM MICRO CURVED BARBELL FOR PTTT ROC	LOT NO 001127	SET/50	500
41	PT-Z-10-TMLB-1.2-10.0-3	STERILE TITANIUM MICRO LABRET FOR PTTT ROC	LOT NO 001127	SET/100	1000
44	PT-Z-10-TMLEJ-1.2-10.0-2.5-CR	STERILE TITANIUM MICRO LABRET WITH JEWELLED BALL FOR PTTT ROC	LOT NO 001127	SET/15	150
45	PT-Z-10-TMLEJ-1.2-10.0-3-CR	STERILE TITANIUM MICRO LABRET WITH JEWELLED BALL FOR PTTT ROC	LOT NO 001127	SET/20	200
47	PT-Z-10-TMLEJ-1.2-8.0-3-CR	STERILE TITANIUM MICRO LABRET WITH JEWELLED BALL FOR PTTT ROC	LOT NO 001127	SET/30	300
		2567015			
3	SV-STERILIZE	SERVICE STERILE	LOT NO 001127		18
		2567976			
3	SV-STERILIZE	SERVICE STERILE	LOT NO 001127		99
					5934

Sterilization:
Temperature: 44.1°C
Humidity: 20.194
Pressures: -2.26Pa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-1-13 13:44

Add ED gas
Temperature: 43.9°C
Humidity: 20.284
Pressures: -4.06Pa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-1-13 13:44

Evacuation
Temperature: 44.2°C
Humidity: 20.194
Pressures: -0.06Pa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-1-13 13:44

Keeping Temp
Temperature: 42.6°C
Humidity: 20.194
Pressures: -1.66Pa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-1-13 12:39

Keep pressure qualified
Temperature: 41.5°C
Humidity: 20.194
Pressures: -60.06Pa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-1-13 12:37

Evacuation
Temperature: 41.5°C
Humidity: 20.194
Pressures: -3.06Pa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-1-13 12:26

Heating-up
Temperature: 21.1°C
Humidity: 20.194
Pressures: -3.06Pa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-1-13 11:42

Setting
Temperature: 45.0°C
Humidity: 20.064
Pressures: -60.06Pa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-1-13 11:42

END

10/1/26
Lot 4129
Air replacement:
Temperature: 45.4°C
Humidity: 20.194
Pressures: -1.66Pa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-1-13 18:12

Air replacement:
Temperature: 45.9°C
Humidity: 20.064
Pressures: -1.06Pa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-1-13 17:52

Air replacement:
Temperature: 45.9°C
Humidity: 20.194
Pressures: -1.66Pa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-1-13 17:31

Air replacement:
Temperature: 45.2°C
Humidity: 20.194
Pressures: -1.06Pa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-1-13 17:11

Air replacement:
Temperature: 45.1°C
Humidity: 20.194
Pressures: -1.06Pa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-1-13 16:51

Sterilization:
Temperature: 44.3°C
Humidity: 20.194
Pressures: -0.06Pa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-1-13 16:45

Sterilization:
Temperature: 44.1°C
Humidity: 20.194
Pressures: -0.06Pa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-1-13 13:40