


Report Test

M/C: Henan Sanqiang Medical Equipment Liability Co. Ltd
 Model: SQ-H220
 Serial No: HNSQ2020050063

LOT No: 001135
 Date: 27 January 2026


Determination of Sterilization Operation Specification

Item Detail	Specification	Setting	Result
Chamber Setting Temperature	40.0-55.5 °C	45.0 °C	
Vacuum pressure	40-60 kPa	60 kPa	
Vacuum Time	50-60 min	60 min	
Humidification			
Humidity	50.0-90.0%	20.2 %	
Sterilization			
Sterilization Temperature	40.0-55.5 °C	43.5 °C	
Gas Injection			
Final pressure	0-10 kPa	4 kPa	
Final temperature of vaporizer	40.0-55.5 °C	44.8 °C	
Exposure			
Temperature at start point of Exposure time period	40.0-55.5 °C	43.3 °C	
Temperature at finish point of Exposure time period	40.0-55.5 °C	46.0 °C	
Exposure Time	120-240 min	180 min	
Exhaust			
Exhaust Time	60-80 min	75 min	
First Vacuum pressure	0-2 kPa	1.5 kPa	
Second Vacuum pressure	0-2 kPa	1.5 kPa	
Third Vacuum pressure	0-2 kPa	1.5 kPa	
Fourth Vacuum pressure	0-2 kPa	1.5 kPa	
Fifth Vacuum pressure	0-2 kPa	1.5 kPa	
Storage			
Storage Time	≥ 48 hour	≥ 48 hour	
Chemical Test			
EO sterilization temperature 40-55°C, Time Sterilization by EO 3hours	Blue	Red	Pass
Biological Test			
EZ Test Biological Indicator	Green	Green	Pass


 (.....)
 Mr. Jutha Liamukda
 Research and Development Engineer



			LOT	Color	BOX	TOTAL / PCS
		217749				
1	Z-BKMBN-1.2-10-0-3	STERILE BLACK PVD SURGICAL STEEL CURVED MICRO BARBELL	LOT NO 001135			10
2	Z-BKMBN-1.2-12-0-3	STERILE BLACK PVD SURGICAL STEEL CURVED MICRO BARBELL	LOT NO 001135			10
3	Z-BKMLB-1.2-10-0-3	STERILE BLACK PVD SURGICAL STEEL MICRO LABRET	LOT NO 001135			10
4	Z-BKMLB-1.2-12-0-3	STERILE BLACK PVD SURGICAL STEEL MICRO LABRET	LOT NO 001135			5
5	Z-BKMLB-1.2-8-0-3	STERILE BLACK PVD SURGICAL STEEL MICRO LABRET	LOT NO 001135			10
6	Z-BL-1.8-10-0-3	STERILE SURGICAL STEEL BARBELL	LOT NO 001135			5
7	Z-BL-1.8-12-0-3	STERILE SURGICAL STEEL BARBELL	LOT NO 001135			5
8	Z-BL-1.8-14-0-3	STERILE SURGICAL STEEL BARBELL	LOT NO 001135			5
9	Z-BNN-1.8-10-0-4/5	STERILE SURGICAL STEEL NAVEL BANANA	LOT NO 001135			10
10	Z-BNN-1.8-10-0-4/6	STERILE SURGICAL STEEL NAVEL BANANA	LOT NO 001135			10
11	Z-GPBNJ-1.8-10-0-5/8-CR	STERILE GOLD PVD COATED SURGICAL STEEL PREMIUM CRYSTAL JEWELLED NAVEL BANANA	LOT NO 001135			5
12	Z-GPBNJ-1.8-12-0-5/8-CR	STERILE GOLD PVD COATED SURGICAL STEEL PREMIUM CRYSTAL JEWELLED NAVEL BANANA	LOT NO 001135			5
13	Z-GPMBN-1.2-10-0-3	STERILE GOLD PVD COATED SURGICAL STEEL CURVED MICRO BARBELL	LOT NO 001135			10
14	Z-GPMBN-1.2-8-0-3	STERILE GOLD PVD COATED SURGICAL STEEL CURVED MICRO BARBELL	LOT NO 001135			10
15	Z-GPTNO-OP-1.0-8-5-2-CR	STERILE GOLD PVD COATED TITANIUM OPTIMA CRYSTAL JEWELLED CURVED NOSE STUD	LOT NO 001135			5
16	Z-INMLBSJ-1.0-8-0-2-CR	STERILE SURGICAL STEEL INTERNALLY THREADED ROUND PRONG SET JEWELLED MICRO LABRET	LOT NO 001135			5
17	Z-INMLBSJ-1.0-8-0-3-CR	STERILE SURGICAL STEEL INTERNALLY THREADED ROUND PRONG SET JEWELLED MICRO LABRET	LOT NO 001135			5
18	Z-INMLBSJ-1.2-10-0-2-CR	STERILE SURGICAL STEEL INTERNALLY THREADED ROUND PRONG SET JEWELLED MICRO LABRET	LOT NO 001135			10
19	Z-INMLBSJ-1.2-8-0-2-CR	STERILE SURGICAL STEEL INTERNALLY THREADED ROUND PRONG SET JEWELLED MICRO LABRET	LOT NO 001135			5
20	Z-INO-0.8-6-5-1.8-AB	STERILE SURGICAL STEEL PREMIUM CRYSTAL JEWELLED CURVED NOSE STUD	LOT NO 001135			5
21	Z-INO-0.8-6-5-1.8-AQ	STERILE SURGICAL STEEL PREMIUM CRYSTAL JEWELLED CURVED NOSE STUD	LOT NO 001135			5
22	Z-INO-0.8-6-5-2.35-BZ	STERILE SURGICAL STEEL PREMIUM CRYSTAL JEWELLED CURVED NOSE STUD	LOT NO 001135			5
23	Z-INO-0.8-6-5-2.35-CA	STERILE SURGICAL STEEL PREMIUM CRYSTAL JEWELLED CURVED NOSE STUD	LOT NO 001135			5
24	Z-INO-0.8-6-5-2.35-CC	STERILE SURGICAL STEEL PREMIUM CRYSTAL JEWELLED CURVED NOSE STUD	LOT NO 001135			5
25	Z-INO-0.8-6-5-2.35-FU	STERILE SURGICAL STEEL PREMIUM CRYSTAL JEWELLED CURVED NOSE STUD	LOT NO 001135			5
26	Z-LB-1.8-10-0-3	STERILE SURGICAL STEEL LABRET	LOT NO 001135			5
27	Z-LB-1.8-12-0-3	STERILE SURGICAL STEEL LABRET	LOT NO 001135			5
28	Z-LB-1.8-8-0-3	STERILE SURGICAL STEEL LABRET	LOT NO 001135			5
29	Z-MBN-1.0-8-0-2.5	STERILE SURGICAL STEEL CURVED MICRO BARBELL	LOT NO 001135			5
30	Z-MBN-1.0-8-0-2	STERILE SURGICAL STEEL CURVED MICRO BARBELL	LOT NO 001135			5
31	Z-MBN-1.0-8-0-3	STERILE SURGICAL STEEL CURVED MICRO BARBELL	LOT NO 001135			5
32	Z-MBN-1.2-10-0-2	STERILE SURGICAL STEEL CURVED MICRO BARBELL	LOT NO 001135			5
33	Z-MLB-1.0-10-0-2	STERILE SURGICAL STEEL MICRO LABRET	LOT NO 001135			25
34	Z-MR-1.0-8-0-2	STERILE SURGICAL STEEL MICRO LABRET	LOT NO 001135			5
35	Z-MLB-1.0-8-0-2	STERILE SURGICAL STEEL MICRO LABRET	LOT NO 001135			5
36	Z-MLB-1.2-10-0-2	STERILE SURGICAL STEEL MICRO LABRET	LOT NO 001135			5
37	Z-MLB-1.2-12-0-2	STERILE SURGICAL STEEL MICRO LABRET	LOT NO 001135			5
38	Z-MLB-1.2-8-0-2	STERILE SURGICAL STEEL MICRO LABRET	LOT NO 001135			5
39	Z-RGBL-1.8-12-0-4	STERILE ROSE GOLD PVD COATED SURGICAL STEEL BARBELL	LOT NO 001135			5
40	Z-RGBL-1.8-14-0-4	STERILE ROSE GOLD PVD COATED SURGICAL STEEL BARBELL	LOT NO 001135			2
41	Z-RGBN-1.2-10-0-3	STERILE ROSE GOLD PVD COATED SURGICAL STEEL CURVED MICRO BARBELL	LOT NO 001135			5
42	Z-RGBN-1.2-8-0-3	STERILE ROSE GOLD PVD COATED SURGICAL STEEL CURVED MICRO BARBELL	LOT NO 001135			5
43	Z-ZGMLB-1.2-10-0-3	STERILE ZIRCON GOLD PVD COATED SURGICAL STEEL MICRO LABRET	LOT NO 001135			5
44	Z-ZGMLB-1.2-8-0-3	STERILE ZIRCON GOLD PVD COATED SURGICAL STEEL MICRO LABRET	LOT NO 001135			6
47	Z-INO-0.8-6-5-1.8-CR	STERILE SURGICAL STEEL PREMIUM CRYSTAL JEWELLED CURVED NOSE STUD	LOT NO 001135			5
		217755				
1	Z-BNDJ-1.8-10-0-4/8-BZ	STERILE SURGICAL STEEL DOUBLE PREMIUM CRYSTALS JEWELLED NAVEL BANANA	LOT NO 001135			1
2	Z-BNDJ-1.8-11-2-5/8-CR	STERILE SURGICAL STEEL DOUBLE PREMIUM CRYSTALS JEWELLED NAVEL BANANA	LOT NO 001135			1
3	Z-L-INO-0.8-5-1.8-CR	STERILE SURGICAL STEEL PREMIUM CRYSTAL JEWELLED LARGE LEFT CURVE NOSE STUD	LOT NO 001135			2
4	Z-L-INO-0.8-5-2.35-CR	STERILE SURGICAL STEEL PREMIUM CRYSTAL JEWELLED LARGE LEFT CURVE NOSE STUD	LOT NO 001135			1
5	Z-L-INO-1.0-8-5-1.8-AB	STERILE SURGICAL STEEL PREMIUM CRYSTAL JEWELLED LARGE LEFT CURVE NOSE STUD	LOT NO 001135			1
6	Z-L-INO-1.0-8-5-2.35-AQ	STERILE SURGICAL STEEL PREMIUM CRYSTAL JEWELLED LARGE LEFT CURVE NOSE STUD	LOT NO 001135			1
7	Z-L-INO-1.0-8-5-2.35-CR	STERILE SURGICAL STEEL PREMIUM CRYSTAL JEWELLED LARGE LEFT CURVE NOSE STUD	LOT NO 001135			1
8	Z-L-INO-1.0-8-5-2.35-RO	STERILE SURGICAL STEEL PREMIUM CRYSTAL JEWELLED LARGE LEFT CURVE NOSE STUD	LOT NO 001135			1
9	Z-MBN-1.2-4-0-2	STERILE SURGICAL STEEL CURVED MICRO BARBELL	LOT NO 001135			1
		217791				
1	Z-DKTN-5-0-5-2	STERILE BLACK PVD COATED STAINLESS STEEL THREADED TUNNCL	LOT NO 001135			1
2	Z-TMBL-1.2-14-0-4-HP	STERILE TITANIUM MICRO BARBELL	LOT NO 001135			2
3	Z-XBNIP2-1.6-16-0-4-CL-AQ	STERILE BIOFLEX VERTICAL HOOD BANANAS WITH SURGICAL STEEL CHARMS - CAN BE CUT TO SIZE	LOT NO 001135			1
4	Z-XCN-1.6-12-0-5/8-CL-CR	STERILE BIOFLEX JEWELLED CUP NAVEL BANANA WITH STEEL BALL	LOT NO 001135			1
5	Z-76TRND.LOP.1.6-10-0-5/8-CR	STERILE ZIRCON GOLD PVD COATED TITANIUM OPTIMA CRYSTAL JEWELLED NAVEL BANANA	LOT NO 001135			1
		217807				
48	Z-BKMBN-1.2-10-0-3	STERILE BLACK PVD SURGICAL STEEL CURVED MICRO BARBELL	LOT NO 001135			3
49	Z-BKMBN-1.2-8-0-3	STERILE BLACK PVD SURGICAL STEEL CURVED MICRO BARBELL	LOT NO 001135			3
47	Z-BKMLB-1.2-10-0-3	STERILE BLACK PVD SURGICAL STEEL MICRO LABRET	LOT NO 001135			5
49	Z-BKMLB-1.2-8-0-3	STERILE BLACK PVD SURGICAL STEEL MICRO LABRET	LOT NO 001135			5
50	Z-GPINO-0.8-5-2.35-CR	STERILE GOLD PVD COATED SURGICAL STEEL PREMIUM CRYSTAL JEWELLED CURVED NOSE STUD	LOT NO 001135			10
51	Z-GPINO-1.0-8-5-2.35-CR	STERILE GOLD PVD COATED SURGICAL STEEL PREMIUM CRYSTAL JEWELLED CURVED NOSE STUD	LOT NO 001135			10
52	Z-GPMBN-1.2-10-0-3	STERILE GOLD PVD COATED SURGICAL STEEL CURVED MICRO BARBELL	LOT NO 001135			10
53	Z-GPMBN-1.2-8-0-3	STERILE GOLD PVD COATED SURGICAL STEEL CURVED MICRO BARBELL	LOT NO 001135			10
54	Z-GPTMBN-1.2-10-0-3	STERILE GOLD PVD COATED TITANIUM CURVED MICRO BARBELL	LOT NO 001135			5
55	Z-GPTMBN-1.2-8-0-3	STERILE GOLD PVD COATED TITANIUM CURVED MICRO BARBELL	LOT NO 001135			5
56	Z-INMLBSJ-1.0-8-0-3-CR	STERILE SURGICAL STEEL INTERNALLY THREADED ROUND PRONG SET JEWELLED MICRO LABRET	LOT NO 001135			10
57	Z-INMLBSJ-1.2-8-0-3-CR	STERILE SURGICAL STEEL INTERNALLY THREADED ROUND PRONG SET JEWELLED MICRO LABRET	LOT NO 001135			10
58	Z-INO-0.8-6-5-2.35-CR	STERILE SURGICAL STEEL PREMIUM CRYSTAL JEWELLED CURVED NOSE STUD	LOT NO 001135			10
59	Z-MBNDJ-1.2-10-0-3-CR	STERILE SURGICAL STEEL DOUBLE JEWELLED CURVED MICRO BARBELL	LOT NO 001135			5
60	Z-MBNDJ-1.2-8-0-3-CR	STERILE SURGICAL STEEL DOUBLE JEWELLED CURVED MICRO BARBELL	LOT NO 001135			5
61	Z-MLB-1.2-6-0-3	STERILE SURGICAL STEEL MICRO LABRET	LOT NO 001135			10
62	Z-MLB-1.2-8-0-3	STERILE SURGICAL STEEL MICRO LABRET	LOT NO 001135			20
63	Z-MLB-1.2-10-0-3-CR	STERILE SURGICAL STEEL JEWELLED MICRO LABRET	LOT NO 001135			10
64	Z-MLB-1.2-6-0-3-CR	STERILE SURGICAL STEEL JEWELLED MICRO LABRET	LOT NO 001135			20
65	Z-MLB-1.2-8-0-3-CR	STERILE SURGICAL STEEL JEWELLED MICRO LABRET	LOT NO 001135			20
66	Z-TBNDJ-CN-1.6-12-0-5/8-HP-CR	STERILE TITANIUM DOUBLE VALUE JEWELLED NAVEL BANANA	LOT NO 001135			5
67	Z-TBNDJ-1.6-10-0-5/8-HP-CR	STERILE TITANIUM INTERNALLY THREADED PREMIUM CRYSTAL JEWELLED NAVEL BANANA	LOT NO 001135			5
68	Z-TMBL-1.2-38-0-3-HP	STERILE TITANIUM MICRO BARBELL	LOT NO 001135			5
69	Z-XBNJ-1.6-12-0-5/8-CL-CR	BIOFLEX JEWELLED NAVEL BANANA WITH SURGICAL STEEL BALL	LOT NO 001135			5
70	Z-ZGMBNDJ-1.2-10-0-3-CR	STERILE ZIRCON GOLD PVD COATED SURGICAL STEEL DOUBLE JEWELLED CURVED MICRO BARBELL	LOT NO 001135			5
71	Z-TJNO-CN-1.0-8-5-2.35-HP-CR	STERILE TITANIUM VALUE CRYSTAL JEWELLED CURVED NOSE STUD	LOT NO 001135			5
72	Z-TSBNJ-CN-1.8-12-0-4/8-HP-CR	STERILE TITANIUM VALUE CRYSTAL JEWELLED MINI NAVEL BANANA	LOT NO 001135			5
		217808				
4	Z-BNDJ-1.8-10-0-4/8-CR	STERILE SURGICAL STEEL DOUBLE PREMIUM CRYSTALS JEWELLED NAVEL BANANA	LOT NO 001135			5
5	Z-BNDJ-1.8-12-0-4/8-CR	STERILE SURGICAL STEEL DOUBLE PREMIUM CRYSTALS JEWELLED NAVEL BANANA	LOT NO 001135			5
		217872				
1	Z-TNSC.M020-1.2-CR	STERILE TITANIUM JEWELLED MICRO ATTACHMENT FOR 1.2MM INTERNALLY THREADED PINS	LOT NO 001135			3
2	Z-TNSC.M026-1.2-CR	STERILE TITANIUM JEWELLED MICRO ATTACHMENT FOR 1.2MM INTERNALLY THREADED PINS	LOT NO 001135			5
3	Z-TNSC.M054-1.2-CR	STERILE TITANIUM JEWELLED MICRO ATTACHMENT FOR 1.2MM INTERNALLY THREADED PINS	LOT NO 001135			5



		217873			
10	Z-BL-1.6-10.0-4	STERILE SURGICAL STEEL BARBELL	LOT NO 001135		3
11	Z-BL-1.6-12.0-4	STERILE SURGICAL STEEL BARBELL	LOT NO 001135		3
12	Z-MLB-1.0-8.0-2	STERILE SURGICAL STEEL MICRO LABRET	LOT NO 001135		3
13	Z-MLB-1.2-10.0-2.5	STERILE SURGICAL STEEL MICRO LABRET	LOT NO 001135		5
14	Z-MLB-1.2-8.0-2	STERILE SURGICAL STEEL MICRO LABRET	LOT NO 001135		4
15	Z-TNC-1.0-6.5-2-HP	STERILE TITANIUM CURVED BALL NOSE STUD	LOT NO 001135		10
		217833			
4	Z-TMLBJ-1.2-8.0-3-HP-CR	STERILE TITANIUM MICRO LABRET WITH JEWELLED BALL	LOT NO 001135		50
		217834			
1	Z-XBL-1.16-50.0-5-BK	STERILE BIOFLEX BARBELL WITH STEEL BALLS	LOT NO 001135		20
		217889			
271	Z-TINSCJ041-1.2-CR	STERILE TITANIUM JEWELLED MICRO ATTACHMENT FOR 1.2MM INTERNALLY THREADED PINS	LOT NO 001135		7
		217895			
1	Z-BKMLBSC45-1.2-8.0	STERILE BLACK PVD COATED SURGICAL STEEL MICRO LABRET WITH ATTACHMENT - SALAMANDER	LOT NO 001135		1
2	Z-BL-1.6-12.0-3	STERILE SURGICAL STEEL BARBELL	LOT NO 001135		4
3	Z-MBLSCCM88-1.2-8.0-3-CR	STERILE SURGICAL STEEL HELIX MICRO BARBELL WITH JEWELLED CHARM - STAR	LOT NO 001135		1
4	Z-MBLSCC430-1.2-8.0-3-CR	STERILE SURGICAL STEEL JEWELLED TRAGUS MICRO BARBELL	LOT NO 001135		1
5	Z-SCNC1-1.6-18.0	STERILE SURGICAL STEEL NIPPLE CLICKER - BAT WINGS	LOT NO 001135		2
6	Z-SCNS44-1.6-10.0	STERILE SURGICAL STEEL NIPPLE BAR - GUN	LOT NO 001135		2
		2569568			
1	ZSNEEDLES-1.2-48.0	STERILIZED STAINLESS STEEL NEEDLE	LOT NO 001135		2306
		2569293			
27	PT-Z-XBL-1.1.6-22.0-5-CL	STERILE BIOFLEX BARBELL WITH STEEL BALLS FOR PTIT ROC	LOT NO 001135		50
28	PT-Z-XBL-1.1.6-40.0-5-CL	STERILE BIOFLEX BARBELL WITH STEEL BALLS FOR PTIT ROC	LOT NO 001135		50
		2568779			
7	TDI-Z-TBN-1.6-10.0-4	STERILIZED TITANIUM CURVED BARBELL FOR TDI	LOT NO 001135		200
10	TDI-Z-TBNL-OP-1.6-10.0-5/8-CR	STERILE TITANIUM JEWELLED NAVEL BANANA FOR TDI	LOT NO 001135		300
21	TDI-Z-TMCB-1.2-12.0-3	STERILE TITANIUM MICRO CIRCULAR BARBELL FOR TDI	LOT NO 001135		100
		2570180			
10	PT-Z-GPTNO-0.8-8.5-2	STERILE GOLD PVD COATED TITANIUM CURVED BALL NOSE STUD	LOT NO 001135		150
11	PT-Z-TBCSH-1.2-8.0-HP	STERILE TITANIUM HINGED SEGMENT RING FOR PTIT ROC	LOT NO 001135		100
		25687978			
7	SV-STERILIZE	SERVICE STERILE	LOT NO 001135		47
		2567982			
4	SV-STERILIZE	SERVICE STERILE	LOT NO 001135		4
		2570256			
1	SV-STERILIZE	SERVICE STERILE	LOT NO 001135		30
					3998

Sterilization:
Temperature: 47.0°C
Humidity: 20.2RH
Pressures: -3.8kPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2026-1-27 13:46

Add ED gas
Temperature: 47.2°C
Humidity: 20.2RH
Pressures: -4.0kPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2026-1-27 13:45

Evacuation
Temperature: 44.3°C
Humidity: 20.2RH
Pressures: -0.0kPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2026-1-27 13:43

Keeping Temp
Temperature: 43.4°C
Humidity: 20.1RH
Pressures: -1.8kPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2026-1-27 12:31

Keep pressure qualified
Temperature: 42.1°C
Humidity: 20.1RH
Pressures: -5.0kPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2026-1-27 12:26

Evacuation
Temperature: 41.7°C
Humidity: 20.2RH
Pressures: -3.0kPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2026-1-27 12:17

Heat the gas
Temperature: 37.0°C
Humidity: 20.1RH
Pressures: -3.0kPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2026-1-27 11:38

Setting
Temperature: 45.0°C
Humidity: 20.1RH
Pressures: -5.0kPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2026-1-27 11:36

20/01/26
DNO
Lot 1135

Air replacement:
Temperature: 45.0°C
Humidity: 20.2RH
Pressures: -1.8kPa
Sterilization time: 3Hour
Air replace rounds: 1
round time: 0min(s)
2026-1-27 15:18

Air replacement:
Temperature: 45.5°C
Humidity: 20.2RH
Pressures: -1.8kPa
Sterilization time: 3Hour
Air replace rounds: 2
round time: 0min(s)
2026-1-27 17:58

Air replacement:
Temperature: 45.8°C
Humidity: 20.2RH
Pressures: -1.8kPa
Sterilization time: 3Hour
Air replace rounds: 3
round time: 0min(s)
2026-1-27 17:58

Air replacement:
Temperature: 45.2°C
Humidity: 20.2RH
Pressures: -1.8kPa
Sterilization time: 3Hour
Air replace rounds: 4
round time: 0min(s)
2026-1-27 17:17

Air replacement:
Temperature: 45.5°C
Humidity: 20.2RH
Pressures: -1.8kPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2026-1-27 16:57

Sterilization:
Temperature: 45.4°C
Humidity: 20.2RH
Pressures: -0.8kPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2026-1-27 16:48

Sterilization:
Temperature: 45.4°C
Humidity: 20.1RH
Pressures: -3.8kPa
Sterilization time: 3Hour
Air replace rounds: 5
round time: 15min(s)
2026-1-27 13:43