


Report Test

M/C: Henan Sanqiang Medical Equipment Liability Co. Ltd
 Model: SQ-H220
 Serial No: HNSQ2020050063

LOT No: 001137
 Date: 29 January 2026

Determination of Sterilization Operation Specification

Item Detail	Specification	Setting	Result
Chamber Setting Temperature	40.0-55.5 °C	45.0 °C	
Vacuum pressure	40-60 kPa	60 kPa	
Vacuum Time	50-60 min	60 min	
Humidification			
Humidity	50.0-90.0%	20.2 %	
Sterilization			
Sterilization Temperature	40.0-55.5 °C	43.5 °C	
Gas Injection			
Final pressure	0-10 kPa	4 kPa	
Final temperature of vaporizer	40.0-55.5 °C	44.8 °C	
Exposure			
Temperature at start point of Exposure time period	40.0-55.5 °C	43.3 °C	
Temperature at finish point of Exposure time period	40.0-55.5 °C	46.0 °C	
Exposure Time	120-240 min	180 min	
Exhaust			
Exhaust Time	60-80 min	75 min	
First Vacuum pressure	0-2 kPa	1.5 kPa	
Second Vacuum pressure	0-2 kPa	1.5 kPa	
Third Vacuum pressure	0-2 kPa	1.5 kPa	
Fourth Vacuum pressure	0-2 kPa	1.5 kPa	
Fifth Vacuum pressure	0-2 kPa	1.5 kPa	
Storage			
Storage Time	≥ 48 hour	≥ 48 hour	
Chemical Test			
EO sterilization temperature 40-55°C, Time Sterilization by EO 3hours	Blue	Red	Pass
Biological Test			
EZ Test Biological Indicator	Green	Green	Pass


 (.....)
 Mr. Jutha Liammukda
 Research and Development Engineer



			LOT	Color	BOX	TOTAL / PCS
			217234			
3	Z-TINDAJD-1.6-2.0-4-HP-CR	STERILE TITANIUM INTERNALLY THREADED DERMAL ANCHOR WITH JEWELLED DISC	LOT NO 001137			25
			217598			
10	Z-TINDAJD-1.6-2.0-3-HP-CR	STERILE TITANIUM INTERNALLY THREADED DERMAL ANCHOR WITH JEWELLED DISC	LOT NO 001137			50
			217921			
9	Z-SPINILB1183-1.2-8.0-CR	GOLD PVD COATED SURGICAL STEEL INTERNAL LABRET FOUR GEM VERTICAL BAR	LOT NO 001137			1
10	Z-SPINILB1183-1.0-8.0	GOLD PVD COATED SURGICAL STEEL INTERNAL MICRO LABRET LIGHTNING BOLT TOP	LOT NO 001137			1
			217980			
1	Z-BNDL-1.6-12.0-5/8-CR	STERILE SURGICAL STEEL DOUBLE JEWELLED NAVEL BANANA	LOT NO 001137			75
3	Z-GPINC-1.0-8.5-2.35-CR	STERILE GOLD PLATED SURGICAL STEEL JEWELLED CURVED NOSE STUD	LOT NO 001137			130
4	Z-SPIMBL-1.2-8.0-3	STERILE GOLD PVD COATED SURGICAL STEEL MICRO BARBELL	LOT NO 001137			225
5	Z-JNO-1.0-8.5-2.35-CR	STERILE SURGICAL STEEL PREMIUM CRYSTAL JEWELLED CURVED NOSE STUD	LOT NO 001137			200
6	Z-MBL-1.2-8.0-3	STERILE SURGICAL STEEL MICRO BARBELL	LOT NO 001137			200
			218017			
3	Z-GPBCSH-1.0-7.0	GOLD PVD COATED SURGICAL STEEL HINGED SEGMENT RING PLAIN POLISHED	LOT NO 001137			47
			218018			
2	Z-GPBCSH-1.0-6.0	GOLD PVD COATED SURGICAL STEEL HINGED SEGMENT RING PLAIN POLISHED	LOT NO 001137			50
3	Z-GPBCSH-1.0-7.0	GOLD PVD COATED SURGICAL STEEL HINGED SEGMENT RING PLAIN POLISHED	LOT NO 001137			3
4	Z-GPBCSH-1.0-8.0	GOLD PVD COATED SURGICAL STEEL HINGED SEGMENT RING PLAIN POLISHED	LOT NO 001137			50
			258343			
12	THD-Z-G18INJ204-1.2-CR	STERILE GOLD PVD COATED SURGICAL STEEL JEWELLED MICRO ATTACHMENT FOR 1.2MM INTERNALLY THREADED PINS	LOT NO 001137			400
			2588718			
12	TDI-Z-TBNJ-OP-1.6-14.0-5/8-CR	STERILE TITANIUM JEWELLED NAVEL BANANA FOR TDI	LOT NO 001137			100
			2570180			
20	PT-Z-XMLB-1.1-12.12.0-3/4B-CL	STERILE BIOFLEX MICRO LABRET WITH STEEL BALL FOR PTIT ROC	LOT NO 001137			100
21	PT-Z-XMLB-1.1-12.10.0-3-CL-CR	STERILE BIOFLEX JEWELLED MICRO LABRET FOR PTIT ROC	LOT NO 001137			80
23	PT-Z10-TMCB-1.2-10.0-3	STERILE TITANIUM MICRO CIRCULAR BARBELL FOR PTIT ROC	LOT NO 001137		SET760	500
24	PT-Z10-TMLB-1.2-8.0-3	STERILE TITANIUM MICRO LABRET FOR PTIT ROC	LOT NO 001137		SET7100	1000
			2570335			
33	THD-Z-JNO-1.0-8.5-2.35-BZ	STERILE SURGICAL STEEL JEWELLED CURVED NOSE STUD FOR THENDRO	LOT NO 001137			150
			2587876			
9	SV-STERILIZE	SERVICE STERILE	LOT NO 001137			58
			2570515			
1	SV-STERILIZE	SERVICE STERILE	LOT NO 001137			1
						2417

Sterilization:
Temperature: 45.3°C
Humidity: 20.18%
Pressures: 4.0kPa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-1-29 18:4

Add 60 gas
Temperature: 45.3°C
Humidity: 20.18%
Pressures: 4.0kPa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-1-29 18:4

Evacuation
Temperature: 44.2°C
Humidity: 20.28%
Pressures: 0.0kPa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-1-29 17:54

Keeping Temp
Temperature: 45.0°C
Humidity: 20.18%
Pressures: 2.0kPa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-1-29 16:54

Keep pressure qualified
Temperature: 41.2°C
Humidity: 20.18%
Pressures: 55.2kPa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-1-29 16:52

Evacuation
Temperature: 41.8°C
Humidity: 20.28%
Pressures: 3.0kPa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-1-29 16:41

Heating up
Temperature: 42.8°C
Humidity: 20.28%
Pressures: 3.0kPa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-1-29 16:37

Settings
Temperature: 45.0°C
Humidity: 20.18%
Pressures: 60.0kPa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-1-29 16:36

END 29/01/26
Lot 1007

Air replacement:
Temperature: 45.0°C
Humidity: 20.18%
Pressures: 1.8kPa
Sterilization time: 3hour
Air replace rounds: 1
round time: 0min(s)
2026-1-29 22:31

Air replacement:
Temperature: 45.4°C
Humidity: 20.18%
Pressures: 1.8kPa
Sterilization time: 3hour
Air replace rounds: 2
round time: 0min(s)
2026-1-29 22:11

Air replacement:
Temperature: 45.0°C
Humidity: 20.28%
Pressures: 1.8kPa
Sterilization time: 3hour
Air replace rounds: 3
round time: 0min(s)
2026-1-29 21:50

Air replacement:
Temperature: 46.4°C
Humidity: 20.28%
Pressures: 1.8kPa
Sterilization time: 3hour
Air replace rounds: 4
round time: 0min(s)
2026-1-29 21:30

Air replacement:
Temperature: 45.8°C
Humidity: 20.18%
Pressures: 1.8kPa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-1-29 21:10

Sterilization:
Temperature: 45.2°C
Humidity: 20.18%
Pressures: 3.8kPa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-1-29 21:5

Sterilization:
Temperature: 45.0°C
Humidity: 20.28%
Pressures: 4.2kPa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-1-29 16:5