


Report Test


M/C: Henan Sanqiang Medical Equipment Liability Co. Ltd
 Model: SQ-H220
 Serial No: HNSQ2020050063

LOT No: 001151
 Date: 19 February 2026

Determination of Sterilization Operation Specification

Item Detail	Specification	Setting	Result
Chamber Setting Temperature	40.0-55.5 °C	45.0 °C	
Vacuum pressure	40-60 kPa	60 kPa	
Vacuum Time	50-60 min	60 min	
Humidification			
Humidity	50.0-90.0%	20.2 %	
Sterilization			
Sterilization Temperature	40.0-55.5 °C	43.5 °C	
Gas Injection			
Final pressure	0-10 kPa	4 kPa	
Final temperature of vaporizer	40.0-55.5 °C	44.8 °C	
Exposure			
Temperature at start point of Exposure time period	40.0-55.5 °C	43.3 °C	
Temperature at finish point of Exposure time period	40.0-55.5 °C	46.0 °C	
Exposure Time	120-240 min	180 min	
Exhaust			
Exhaust Time	60-80 min	75 min	
First Vacuum pressure	0-2 kPa	1.5 kPa	
Second Vacuum pressure	0-2 kPa	1.5 kPa	
Third Vacuum pressure	0-2 kPa	1.5 kPa	
Fourth Vacuum pressure	0-2 kPa	1.5 kPa	
Fifth Vacuum pressure	0-2 kPa	1.5 kPa	
Storage			
Storage Time	≥ 48 hour	≥ 48 hour	
Chemical Test			
EO sterilization temperature 40-55°C, Time Sterilization by EO 3hours	Blue	Red	Pass
Biological Test			
EZ Test Biological Indicator	Green	Green	Pass


 (.....)
 Mr. Jutha Liamukda
 Research and Development Engineer



FM-ISO0-009 Rev.02 (05/11/2020)

			LOT	Color	BOX	TOTAL / PCS
		218240				
2	Z-XBNNJ-1-1.6-12 0.5/8-CL-AB	BIOFLEX JEWELLED NAVAL BANANA WITH SURGICAL STEEL BALL	LOT NO 001151			100
		218553				
1	Z-BL-1 6-26 0-4	STERILE SURGICAL STEEL BARBELL	LOT NO 001151			10
2	Z-BL-1 6-50 0-4	STERILE SURGICAL STEEL BARBELL	LOT NO 001151			2
3	Z-BNJ-1-6-14 0.5/8-CR	316L STERILE SURGICAL STEEL NAVAL BANANA CURVED BARBELL SINGLE BEZEL-SET GEM	LOT NO 001151			8
4	Z-CNOD-S-0.8-15 0-1.5	STERILE SURGICAL STEEL STRAIGHT DISC NOSE STUD 15MM	LOT NO 001151			10
5	Z-SES1-2-5-CR	STERILE SURGICAL STEEL ROUND PRONG SET JEWELLED EAR STUDS PAIR	LOT NO 001151			8
		218563				
1	Z-ZGMLB-1-2-8 0-3-CR	STERILE ZIRCON GOLD PVD COATED SURGICAL STEEL JEWELLED MICRO LABRET	LOT NO 001151			4
2	Z-ZGMLB-1-2-10 0-3-CR	STERILE ZIRCON GOLD PVD SURGICAL STEEL JEWELLED MICRO LABRET	LOT NO 001151			4
3	Z-INMLBSJ-1-2-10 0-2-CR	STERILE SURGICAL STEEL INTERNALLY THREADED ROUND PRONG SET JEWELLED MICRO LABRET	LOT NO 001151			4
4	Z-BL-1 6-10 0-3	STERILE SURGICAL STEEL BARBELL	LOT NO 001151			3
5	Z-BNDJ-1 6-12 0-4/8-CR	STERILE SURGICAL STEEL DOUBLE PREMIUM CRYSTALS JEWELLED NAVAL BANANA	LOT NO 001151			3
6	Z-ONR-1 0-10 0-2.5	STERILE SURGICAL STEEL OPEN NOSE RING	LOT NO 001151			2
7	Z-INMLBJD-1 2-10 0-2-CR	STERILE SURGICAL STEEL INTERNALLY THREADED LABRET WITH JEWELLED DISC	LOT NO 001151			8
		218558				
7	Z-INMLBSJ-1-0-8 0-2.5-CR	STERILE SURGICAL STEEL INTERNALLY THREADED ROUND PRONG SET JEWELLED MICRO LABRET	LOT NO 001151			12
		218559				
2	Z-INMLBSJ-1-0-8 0-2.5-CR	STERILE SURGICAL STEEL INTERNALLY THREADED ROUND PRONG SET JEWELLED MICRO LABRET	LOT NO 001151			18
		2569100				
1	ZSNEEDLES-1 0-48 0	STERILIZED STAINLESS STEEL NEEDLE	LOT NO 001151			975
		2670335				
50	THD-Z-TINMLB-1.2-10-0-3	STERILE TITANIUM INTERNALLY THREADED MICRO LABRET	LOT NO 001151			2500
		2571428				
1	TDI-Z-TINMLB-1.2-10 0-3	STERILE TITANIUM INTERNALLY THREADED MICRO LABRET FOR TDI	LOT NO 001151			300
5	TDI-Z-TINMLBJ-1.2-10 0-2-CR	STERILE TITANIUM INTERNALLY THREADED JEWELLED MICRO LABRET FOR TDI	LOT NO 001151			300
6	TDI-Z-TINMLBJ-1.2-10 0-3-CR	STERILE TITANIUM INTERNALLY THREADED JEWELLED MICRO LABRET FOR TDI	LOT NO 001151			200
		2572104				
1	ZSNEEDLES-2 0-48 0	STERILIZED STAINLESS STEEL NEEDLE	LOT NO 001151			1164
		2572696				
1	ZSNEEDLES-1 0-48 0	STERILIZED STAINLESS STEEL NEEDLE	LOT NO 001151			200
		2572709				
4	TDI-Z-TINMLB-1 2-6 0-3	STERILE TITANIUM INTERNALLY THREADED MICRO LABRET FOR TDI	LOT NO 001151			50
						5903

Sterilization:
Temperature: 44.4°C
Humidity: 20.1RH
Pressures: -0.0KPa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-2-19 12:41

Add E9 gas
Temperature: 44.1°C
Humidity: 20.1RH
Pressures: -4.0KPa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-2-19 12:41

Evacuation
Temperature: 44.3°C
Humidity: 20.1RH
Pressures: -0.0KPa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-2-19 12:38

Keeping Temp
Temperature: 42.7°C
Humidity: 20.1RH
Pressures: -1.0KPa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-2-19 11:36

Keep pressure qualified
Temperature: 41.4°C
Humidity: 20.1RH
Pressures: -59.0KPa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-2-19 11:34

Evacuation
Temperature: 41.5°C
Humidity: 20.1RH
Pressures: -3.0KPa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-2-19 11:22

Heating-up:
Temperature: 22.8°C
Humidity: 20.1RH
Pressures: -3.0KPa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-2-19 10:41

Setting:
Temperature: 45.0°C
Humidity: 55.0RH
Pressures: -60.0KPa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-2-19 10:40

19/2/26
LOT 1151

END

Air replacement:
Temperature: 45.0°C
Humidity: 20.1RH
Pressures: -1.0KPa
Sterilization time: 3hour
Air replace rounds: 1
round time: 0min(s)
2026-2-19 17:9

Air replacement:
Temperature: 45.6°C
Humidity: 20.1RH
Pressures: -1.0KPa
Sterilization time: 3hour
Air replace rounds: 2
round time: 0min(s)
2026-2-19 16:49

Air replacement:
Temperature: 46.2°C
Humidity: 20.1RH
Pressures: -1.0KPa
Sterilization time: 3hour
Air replace rounds: 5
round time: 0min(s)
2026-2-19 16:29

Air replacement:
Temperature: 45.7°C
Humidity: 20.1RH
Pressures: -1.0KPa
Sterilization time: 3hour
Air replace rounds: 4
round time: 0min(s)
2026-2-19 16:8

Air replacement:
Temperature: 44.9°C
Humidity: 20.1RH
Pressures: -2.0KPa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-2-19 15:40

Sterilization:
Temperature: 44.0°C
Humidity: 20.2RH
Pressures: -0.0KPa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-2-19 15:41

Sterilization:
Temperature: 44.4°C
Humidity: 20.1RH
Pressures: -0.0KPa
Sterilization time: 3hour
Air replace rounds: 5
round time: 15min(s)
2026-2-19 12:43

