


Report Test

M/C: Henan Sanqiang Medical Equipment Liability Co. Ltd
 Model: SQ-H220
 Serial No: HNSQ2020050063

LOT No: 001156
 Date: 27 February 2026

Determination of Sterilization Operation Specification

Item Detail	Specification	Setting	Result
Chamber Setting Temperature	40.0-55.5 °C	45.0 °C	
Vacuum pressure	40-60 kPa	60 kPa	
Vacuum Time	50-60 min	60 min	
Humidification			
Humidity	50.0-90.0%	20.2 %	
Sterilization			
Sterilization Temperature	40.0-55.5 °C	43.5 °C	
Gas Injection			
Final pressure	0-10 kPa	4 kPa	
Final temperature of vaporizer	40.0-55.5 °C	44.8 °C	
Exposure			
Temperature at start point of Exposure time period	40.0-55.5 °C	43.3 °C	
Temperature at finish point of Exposure time period	40.0-55.5 °C	46.0 °C	
Exposure Time	120-240 min	180 min	
Exhaust			
Exhaust Time	60-80 min	75 min	
First Vacuum pressure	0-2 kPa	1.5 kPa	
Second Vacuum pressure	0-2 kPa	1.5 kPa	
Third Vacuum pressure	0-2 kPa	1.5 kPa	
Fourth Vacuum pressure	0-2 kPa	1.5 kPa	
Fifth Vacuum pressure	0-2 kPa	1.5 kPa	
Storage			
Storage Time	≥ 48 hour	≥ 48 hour	
Chemical Test			
EO sterilization temperature 40-55°C, Time Sterilization by EO 3hours	Blue	Red	Pass
Biological Test			
EZ Test Biological Indicator	Green	Green	Pass


 (.....)
 Mr. Jutha Liamukda
 Research and Development Engineer

			LOT	Color	BOX	TOTAL / PCS
			218804			
7	Z-MLB-1.2-10.0-2.5	STERILE SURGICAL STEEL MICRO LABRET	LOT NO 001156			10
9	Z-MLBJ-1.2-10.0-2.5-CR	316L SURGICAL STEEL LABRET BALL END SINGLE CLEAR GEM FLAT BACK	LOT NO 001156			10
10	Z-MLBJ-1.2-10.0-3-CR	316L SURGICAL STEEL LABRET BALL END SINGLE CLEAR GEM FLAT BACK	LOT NO 001156			10
11	Z-MLBJ-1.2-8.0-2.5-CR	316L SURGICAL STEEL LABRET BALL END SINGLE CLEAR GEM FLAT BACK	LOT NO 001156			10
12	Z-MLBJ-1.2-8.0-3-CR	316L SURGICAL STEEL LABRET BALL END SINGLE CLEAR GEM FLAT BACK	LOT NO 001156			10
15	Z-ZGMLBJ-1.2-10.0-3-CR	STERILE ZIRCON GOLD PVD SURGICAL STEEL JEWELED MICRO LABRET	LOT NO 001156			10
16	Z-ZGMLBJ-1.2-8.0-3-CR	STERILE ZIRCON GOLD PVD SURGICAL STEEL JEWELED MICRO LABRET	LOT NO 001156			10
			218833			
4	Z-BL-1.6-18.0-4	STERILE SURGICAL STEEL BARBELL	LOT NO 001156			1
5	Z-BNJ-1.6-12.0-5-8-CR	316L STERILE SURGICAL STEEL NAVEL BANANA CURVED BARBELL SINGLE BEZEL-SET GEM	LOT NO 001156			3
6	Z-BNN-1.6-12.0-4-6	STERILE SURGICAL STEEL NAVEL BANANA	LOT NO 001156			1
7	Z-JBPACX8-1.8-3-CR	PACK OF 8 STERILE SURGICAL STEEL PREMIUM CRYSTAL JEWELED BALLS	LOT NO 001156			2
8	Z-MBN-1.2-10.0-2	STERILE SURGICAL STEEL CURVED MICRO BARBELL	LOT NO 001156			2
9	Z-MCB-1.2-10.0-2	STERILE SURGICAL STEEL MICRO CIRCULAR BARBELL	LOT NO 001156			3
10	Z-MLB-1.2-10.0-2	STERILE SURGICAL STEEL MICRO LABRET	LOT NO 001156			3
11	Z-MSC101-1.2-CR	STERILE SURGICAL STEEL MICRO THREADED JEWELED FLOWER ATTACHMENT	LOT NO 001156			11
12	Z-SES1-3-CR	STERILE SURGICAL STEEL ROUND PRONG SET JEWELED EAR STUDS PAIR	LOT NO 001156			4
13	Z-TINMLBJD-1.0-8.0-2-HP-CR	STERILE TITANIUM INTERNALLY THREADED JEWELED DISC MICRO LABRET	LOT NO 001156			5
14	Z-TJNO-0.8-6.5-1.8-HP-CR	STERILE TITANIUM PREMIUM CRYSTAL JEWELED CURVED NOSE STUD	LOT NO 001156			5
15	Z-TST-1.6-16.0-4-HP	STERILE TITANIUM 90 DEGREE STAPLE BARBELL	LOT NO 001156			4
			218834			
1	Z-TINDAJD-1.6-2.5-3.5-HP-CR	STERILE TITANIUM INTERNALLY THREADED DERMAL ANCHOR WITH JEWELED DISC	LOT NO 001156			2
			218844			
1	Z-INJBN-1.2-10.0-3	STERILE SURGICAL STEEL INTERNALLY THREADED CURVED MICRO BARBELL	LOT NO 001156			5
2	Z-INJBN-1.2-8.0-3	STERILE SURGICAL STEEL INTERNALLY THREADED CURVED MICRO BARBELL	LOT NO 001156			5
3	Z-INMLBJD-1.0-8.0-2.5-CR	STERILE SURGICAL STEEL INTERNALLY THREADED LABRET WITH JEWELED DISC	LOT NO 001156			5
4	Z-INMLBJD-1.2-10.0-3-CR	STERILE SURGICAL STEEL INTERNALLY THREADED LABRET WITH JEWELED DISC	LOT NO 001156			10
5	Z-INSNBJ-1.6-12.0-4-6-CR	STERILE SURGICAL STEEL INTERNALLY THREADED PREMIUM CRYSTAL JEWELED MINI NAVEL BANANA	LOT NO 001156			10
6	Z-MBL-1.2-10.0-3	STERILE SURGICAL STEEL MICRO BARBELL	LOT NO 001156			5
8	Z-INMLBJD-1.2-10.0-3-CR	STERILE SURGICAL STEEL INTERNALLY THREADED LABRET WITH JEWELED DISC	LOT NO 001156			10
			218846			
1	Z-MBLSJ1-1.2-8.0-3-CR	STERILE SURGICAL STEEL ROUND PRONG SET JEWELED TRAGUS MICRO BARBELL	LOT NO 001156			75
2	Z-TBL-1.6-16.0-5-HP	STERILE TITANIUM BARBELL	LOT NO 001156			50
			2570610			
6	SV-STERILIZE	SERVICE STERILE	LOT NO 001156			175
						466



Sterilization:
Temperature: 45.0K
Humidity:20.1RH
Pressures:-3.0KPa
Sterilization time:3Hour
Air replace rounds:5
round time:15min(s)
2026-2-27 13:44

Add EO gas
Temperature: 45.0K
Humidity:20.1RH
Pressures:-4.0KPa
Sterilization time:3Hour
Air replace rounds:5
round time:15min(s)
2026-2-27 13:44

Evacuation
Temperature: 44.2K
Humidity:20.2RH
Pressures:-0.0KPa
Sterilization time:3Hour
Air replace rounds:5
round time:15min(s)
2026-2-27 13:38

Keeping Temp
Temperature: 43.0K
Humidity:20.1RH
Pressures:-1.0KPa
Sterilization time:3Hour
Air replace rounds:5
round time:15min(s)
2026-2-27 12:38

Keep pressure qualified
Temperature: 41.5K
Humidity:20.1RH
Pressures:-58.0KPa
Sterilization time:3Hour
Air replace rounds:5
round time:15min(s)
2026-2-27 12:38

Evacuation
Temperature: 41.5K
Humidity:20.1RH
Pressures:-0.0KPa
Sterilization time:3Hour
Air replace rounds:5
round time:15min(s)
2026-2-27 12:24

Heating-up:
Temperature: 22.2K
Humidity:20.1RH
Pressures:-3.0KPa
Sterilization time:3Hour
Air replace rounds:5
round time:15min(s)
2026-2-27 11:41

Setting:
Temperature: 45.0K
Humidity:55.0RH
Pressures:-80.0KPa
Sterilization time:3Hour
Air replace rounds:5
round time:15min(s)
2026-2-27 11:40

END
27/02/26
Lot 115b

Air replacement:
Temperature: 45.0K
Humidity:20.1RH
Pressures:-1.0KPa
Sterilization time:3Hour
Air replace rounds:1
round time:0min(s)
2026-2-27 10:12

Air replacement:
Temperature: 45.7K
Humidity:20.1RH
Pressures:-1.0KPa
Sterilization time:3Hour
Air replace rounds:2
round time:0min(s)
2026-2-27 17:52

Air replacement:
Temperature: 45.0K
Humidity:20.1RH
Pressures:-1.0KPa
Sterilization time:3Hour
Air replace rounds:3
round time:0min(s)
2026-2-27 17:31

Air replacement:
Temperature: 45.2K
Humidity:20.1RH
Pressures:-1.0KPa
Sterilization time:3Hour
Air replace rounds:4
round time:0min(s)
2026-2-27 17:11

Air replacement:
Temperature: 44.5K
Humidity:20.2RH
Pressures:-1.0KPa
Sterilization time:3Hour
Air replace rounds:5
round time:15min(s)
2026-2-27 16:50

Sterilization:
Temperature: 44.5K
Humidity:20.1RH
Pressures:-0.0KPa
Sterilization time:3Hour
Air replace rounds:5
round time:15min(s)
2026-2-27 16:44

Sterilization:
Temperature: 44.4K
Humidity:20.2RH
Pressures:-0.0KPa
Sterilization time:3Hour
Air replace rounds:5
round time:15min(s)
2026-2-27 13:46