


Report Test

M/C: Henan Sanqiang Medical Equipment Liability Co. Ltd
 Model: SQ-H220
 Serial No: HNSQ2020050063

LOT No: 001167
 Date: 16 March 2026

Determination of Sterilization Operation Specification

Item Detail	Specification	Setting	Result
Chamber Setting Temperature	40.0-55.5 °C	45.0 °C	
Vacuum pressure	40-60 kPa	60 kPa	
Vacuum Time	50-60 min	60 min	
Humidification			
Humidity	50.0-90.0%	20.2 %	
Sterilization			
Sterilization Temperature	40.0-55.5 °C	43.5 °C	
Gas Injection			
Final pressure	0-10 kPa	4 kPa	
Final temperature of vaporizer	40.0-55.5 °C	44.8 °C	
Exposure			
Temperature at start point of Exposure time period	40.0-55.5 °C	43.3 °C	
Temperature at finish point of Exposure time period	40.0-55.5 °C	46.0 °C	
Exposure Time	120-240 min	180 min	
Exhaust			
Exhaust Time	60-80 min	75 min	
First Vacuum pressure	0-2 kPa	1.5 kPa	
Second Vacuum pressure	0-2 kPa	1.5 kPa	
Third Vacuum pressure	0-2 kPa	1.5 kPa	
Fourth Vacuum pressure	0-2 kPa	1.5 kPa	
Fifth Vacuum pressure	0-2 kPa	1.5 kPa	
Storage			
Storage Time	≥ 48 hour	≥ 48 hour	
Chemical Test			
EO sterilization temperature 40-55°C, Time Sterilization by EO 3hours	Blue	Red	Pass
Biological Test			
EZ Test Biological Indicator	Green	Green	Pass


 (.....)
 Mr. Jutha Liammukda
 Research and Development Engineer



FM-ISO0-009 Rev.02 (05/11/2020)

			LOT	Color	BOX	TOTAL / PCS
		219178				
29	Z-6PBCSH-1 0-8 0	GOLD PVD COATED SURGICAL STEEL HINGED SEGMENT RING PLAIN POLISHED	LOT NO 1167			10
30	Z-INMLBSJ1-1.2-8 0-3-CR	STERILE SURGICAL STEEL INTERNALLY THREADED ROUND PRONG SET JEWELED MICRO LABRET	LOT NO 1167			10
31	Z-INMLBSJ1-1.2-8 0-3-JE	STERILE SURGICAL STEEL INTERNALLY THREADED ROUND PRONG SET JEWELED MICRO LABRET	LOT NO 1167			2
32	Z-INMLBSJ1-1.2-8 0-3-RO	STERILE SURGICAL STEEL INTERNALLY THREADED ROUND PRONG SET JEWELED MICRO LABRET	LOT NO 1167			6
33	Z-INMLBSJ1-1.2-8 0-3-CR	STERILE SURGICAL STEEL INTERNALLY THREADED ROUND PRONG SET JEWELED MICRO LABRET	LOT NO 1167			4
34	Z-MNSC24-1.2-8 0-3-CR	316L SURGICAL STEEL CURVED MICRO BARBELL TRIPLE MARQUISE GEM CLUSTER	LOT NO 1167			4
35	Z-DPU-1.5	STERILE DERMAL PUNCH	LOT NO 1167			20
		219179				
1	Z-8KBCSH-1 0-10 0	STERILE BLACK PVD COATED SURGICAL STEEL HINGED SEGMENT RING	LOT NO 1167			10
2	Z-8KBCSH-1 2-10 0	STERILE BLACK PVD COATED SURGICAL STEEL HINGED SEGMENT RING	LOT NO 1167			10
3	Z-INMLBSJ1-1.2-8 0-3-AM	STERILE SURGICAL STEEL INTERNALLY THREADED ROUND PRONG SET JEWELED MICRO LABRET	LOT NO 1167			8
4	Z-INMLBSJ1-1.2-8 0-3-AQ	STERILE SURGICAL STEEL INTERNALLY THREADED ROUND PRONG SET JEWELED MICRO LABRET	LOT NO 1167			1
5	Z-INMLBSJ1-1.2-8 0-3-SA	STERILE SURGICAL STEEL INTERNALLY THREADED ROUND PRONG SET JEWELED MICRO LABRET	LOT NO 1167			4
6	Z-INMLBSJ1-1.2-8 0-4-CR	STERILE SURGICAL STEEL INTERNALLY THREADED ROUND PRONG SET JEWELED MICRO LABRET	LOT NO 1167			10
		219202				
1	Z-TBL-1.8-14 0-5-HP	STERILE TITANIUM BARBELL	LOT NO 1167			50
		219206				
18	Z-MNSC24-1.2-8 0-3-CR	316L SURGICAL STEEL CURVED MICRO BARBELL TRIPLE MARQUISE GEM CLUSTER	LOT NO 1167			2
		219186				
1	Z-MLB-1.2-10 0-2	STERILE SURGICAL STEEL MICRO LABRET	LOT NO 1167			4
		219187				
2	Z-INMLBJD-1.2-10 0-2.5-CR	STERILE SURGICAL STEEL INTERNALLY THREADED LABRET WITH JEWELED DISC	LOT NO 1167			2
		2574302				
1	THD-Z-BKBC-1 2-10 0-4	STERILE BLACK PVD COATED SURGICAL STEEL BALL CLOSURE RING FOR THENDRO	LOT NO 1167			200
6	THD-Z-BNDJ-1.8-10.0-5B-CR	STERILE SURGICAL STEEL DOUBLE JEWELED NAVEL BAWANA FOR THENDRO	LOT NO 1167			200
20	THD-Z-MCB-1 2-10 0-3	STERILE SURGICAL STEEL MICRO CIRCULAR BARBELL FOR THENDRO	LOT NO 1167			250
23	THD-Z-RGINJB-1.8-4-CR	STERILE ROSE GOLD PVD COATED SURGICAL STEEL PREMIUM CRYSTAL JEWELED BALL FOR 1.8MM INTERNALLY THREADED PIN	LOT NO 1167			200
		2574418				
1	THD-Z-TBL-1.8-10 0-4	STERILE TITANIUM BARBELL FOR THENDRO	LOT NO 1167			200
		2574885				
1	PT-2-GPTMBL-1.2-10 0-3	STERILE GOLD PVD COATED TITANIUM MICRO BARBELL	LOT NO 1167			50
3	PT-2-GPTMCB-1 2-8 0-3	STERILE GOLD PVD COATED TITANIUM MICRO CIRCULAR BARBELL	LOT NO 1167			50
4	PT-2-GPTMLB-1.2-8 0-3	STERILE GOLD PVD COATED TITANIUM MICRO CIRCULAR BARBELL	LOT NO 1167			300
7	PT-2-TBL-1.8-14 0-5	STERILIZED TITANIUM BARBELL FOR PTTI ROC	LOT NO 1167			150
10	PT-2-TMCB-1.2-8 0-3	STERILE TITANIUM MICRO CIRCULAR BARBELL FOR PTTI ROC	LOT NO 1167			100
11	PT-2-TMLB-1 2-10 0-3	STERILE TITANIUM MICRO LABRET FOR PTTI ROC	LOT NO 1167			100
12	PT-2-TMLB-1 2-8 0-3	STERILE TITANIUM MICRO LABRET FOR PTTI ROC	LOT NO 1167			100
17	PT-2-XMLB-1.1 2-10.0-3/4B-CL	STERILE BIOLEX MICRO LABRET WITH STEEL BALL FOR PTTI ROC	LOT NO 1167			50
19	PT-210-TMLB-1 2-8 0-3	STERILE TITANIUM MICRO LABRET FOR PTTI ROC	LOT NO 1167		SET/100	1000
						3165

Sterilization:
 Temperature: 44.0°C
 Humidity: 20.18%
 Pressure: -0.08Pa
 Sterilization time: 20min
 Air replace rounds: 5
 round time: 15min(s)
 2025-3-16 13:46

Add F0 gas
 Temperature: 43.1°C
 Humidity: 20.18%
 Pressure: -4.08Pa
 Sterilization time: 20min
 Air replace rounds: 5
 round time: 15min(s)
 2025-3-16 13:46

Evacuation
 Temperature: 44.2°C
 Humidity: 20.18%
 Pressure: -6.08Pa
 Sterilization time: 20min
 Air replace rounds: 5
 round time: 15min(s)
 2025-3-16 13:41

Keeping temp
 Temperature: 47.0°C
 Humidity: 20.18%
 Pressure: -1.08Pa
 Sterilization time: 20min
 Air replace rounds: 5
 round time: 15min(s)
 2025-3-16 13:41

Keep pressure qualified
 Temperature: 41.5°C
 Humidity: 20.18%
 Pressure: -6.08Pa
 Sterilization time: 20min
 Air replace rounds: 5
 round time: 15min(s)
 2025-3-16 13:39

Evacuation
 Temperature: 43.4°C
 Humidity: 20.18%
 Pressure: -5.08Pa
 Sterilization time: 20min
 Air replace rounds: 5
 round time: 15min(s)
 2025-3-16 13:39

Heating-up
 Temperature: 27.0°C
 Humidity: 20.18%
 Pressure: -3.08Pa
 Sterilization time: 20min
 Air replace rounds: 5
 round time: 15min(s)
 2025-3-16 11:44

Sealing
 Temperature: 45.0°C
 Humidity: 20.18%
 Pressure: -6.08Pa
 Sterilization time: 20min
 Air replace rounds: 5
 round time: 15min(s)
 2025-3-16 11:44

25/03/25
 END
 Lot 1167

Air replacement:
 Temperature: 45.0°C
 Humidity: 20.18%
 Pressure: -1.08Pa
 Sterilization time: 20min
 Air replace rounds: 1
 round time: 6min(s)
 2025-3-16 16:11

Air replacement:
 Temperature: 45.0°C
 Humidity: 20.18%
 Pressure: -1.08Pa
 Sterilization time: 20min
 Air replace rounds: 2
 round time: 6min(s)
 2025-3-16 17:04

Air replacement:
 Temperature: 45.0°C
 Humidity: 20.20%
 Pressure: -1.08Pa
 Sterilization time: 20min
 Air replace rounds: 5
 round time: 6min(s)
 2025-3-16 17:34

Air replacement:
 Temperature: 45.9°C
 Humidity: 20.18%
 Pressure: -1.08Pa
 Sterilization time: 20min
 Air replace rounds: 4
 round time: 6min(s)
 2025-3-16 17:13

Air replacement:
 Temperature: 45.0°C
 Humidity: 20.18%
 Pressure: -1.08Pa
 Sterilization time: 20min
 Air replace rounds: 5
 round time: 15min(s)
 2025-3-16 16:42

Sterilization:
 Temperature: 44.5°C
 Humidity: 20.18%
 Pressure: -6.08Pa
 Sterilization time: 20min
 Air replace rounds: 5
 round time: 15min(s)
 2025-3-16 16:46

Sterilization:
 Temperature: 44.3°C
 Humidity: 20.20%
 Pressure: -0.08Pa
 Sterilization time: 20min
 Air replace rounds: 5
 round time: 15min(s)
 2025-3-16 15:40